

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Dep Johnson AGENCY LPCSO  
DATE 12-17-15 TIME 1106  
INTOXILYZER SERIAL 80-000917 LOCATION QUARTZSITE

- ~~1~~ 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- ~~2~~ 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ~~3~~ 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ~~4~~ 4. Air blank completed
- ~~5~~ 5. Calibration check completed. Test results 0.097 AC.
- ~~6~~ 6. Air blank completed
- ~~7~~ 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE \_\_\_\_\_

INTOXILYZER 8000

Location: QUARTZSITE PD

Serial Number: 80300917

Core Version: 8105.48

12/17/2015

11:08:03

Standard Lot#: 07014100A1

Last Changed By: R. ROSADO #557

QAS: M. JOHNSON #531

LA PAZ: CSO

Operator: M. JOHNSON #531

LA PAZ: CSO

Test	g/210L	Time
Air Blank	0.000	11:08:35
0.100 Cal Check	0.097	11:08:56
Air Blank	0.000	11:09:26

Preventative Maintenance  
Performed -- Timer Reset