

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST WALKER, J # 590 AGENCY LPCSO  
DATE 4-10-16 TIME 1820  
INTOXILYZER SERIAL # 80-001196 LOCATION EHRENBURG SUBSTATION

- (C) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 096 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 590

INTOXILYZER 8000

Location: EHRENBERG SUBSTATION

Serial Number: 80-001196

Core Version: 8105.48

04/10/2016

18:20:34

-----  
Standard Lot#: 08915100A2

Last Changed By: E. SCHNEIDER #431  
-----

QAS: J. WALKER #590

LA PAZ CSO  
-----

Operator: J. WALKER #590

LA PAZ CSO  
-----

Test	g/210L	Time
Air Blank	0.000	18:21:23
0.100 Cal Check	0.096	18:21:44
Air Blank	0.000	18:22:13

Preventative Maintenance  
Performed -- Timer Reset