

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Dep Johnson AGENCY LPCSO
DATE 12-16-15 TIME 1556
INTOXILYZER SERIAL # 80-003861 LOCATION BOATHOUSE

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed
5. Calibration check completed. Test results 0.100 AC.
6. Air blank completed
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000

Location: LPCSO BOATHOUSE

Serial Number: 80-003861

Core Version: 8105.48

12/16/2015

15:57:42

Standard Lot#: 14014100A1

Last Changed By: E. SCHNEIDER #431

QAS: M. JOHNSON #531

LA PAZ CSO

Operator: M. JOHNSON #531

LA PAZ CSO

Test g/210L Time

Air Blank 0.000 15:58:13
0.100 Cal Check 0.100 15:58:34
Air Blank 0.000 15:59:04

Preventative Maintenance
Performed -- Timer Reset